




Bariatric Stetcher (Stryker MX Pro)



Escambia County, Florida - ALS/BLS Medical Protocol



Standard Features:

- Transport at load height capability
- Positive action height adjustment
- Easy to use release handle design
- Color-coded controls
- Dampened action during hot drops



Approved: 10/01/2011
Version: 3.1

Page 1 of 4

Approved by:

Charles Neal, D.O. Medical Director



Bariatric Stetcher (Stryker MX Pro)



Escambia County, Florida - ALS/BLS Medical Protocol

- Lap belt extensions
- High visibility powder-coated frame
- Light-weight, rugged aluminum construction
- Scientifically-optimized lift bar design
- Lower lifting bar
- Seven height positions
- Integrated bumper system
- Lift-capable safety bar
- Perforated litter surface
- One-hand release breakaway head section with safety bar
- Floor-mounted safety hook
- One-hand release fold-down siderails
- One-hand release, infinite positioning, lift assist pneumatic backrest
- Oversized wheels with sealed caster and wheel bearings
- Reflective labeling
- Sealed flat mattress
- Shock position
- Two lap belts and one four-point shoulder restraint

Optional Features

- Heavy duty two or three-stage IV poles (patient right or left)
- Base storage tray
- Non-pocketed head-end storage
- Height limit kit *
- Permanent or removable oxygen bottle holders (foot end)
- Rigid head and foot end push/pull handle



Approved: 10/01/2011
Version: 3.1

Page 2 of 4

Approved by:


Charles Neal, D.O. Medical Director



Bariatric Stetcher (Stryker MX Pro)



Escambia County, Florida - ALS/BLS Medical Protocol

- Single or dual wheel lock
- X-frame guards
- Patient security straps
- Side lift handles
- Tow package
- Transfer-Flat (9' x 5', 1,600 lb capacity)

Specifications:

Model Number	6083
--------------	------

Height ¹

Position 1	13.5"	(34 cm)
Position 2	24.5"	(62 cm)
Position 3	27.5"	(70 cm)
Position 4	30"	(76 cm)
Position 5	32.5"	(83 cm)
Position 6	34.5"	(88 cm)
Position 7	37"	(94 cm)

Length

Standard	80"	(203 cm)
Minimum	62"	(157cm)

Width	29"	(74 cm)
-------	-----	---------

Weight ²	111 lbs	(50kg)
---------------------	---------	--------

Wheels



Approved: 10/01/2011
Version: 3.1

Page 3 of 4

Approved by:


Charles Neal, D.O. Medical Director



Bariatric Stetcher (Stryker MX Pro)



Escambia County, Florida - ALS/BLS Medical Protocol

Diameter	6"	(15cm)
Width	2"	(5cm)
Articulation		
Backrest	0° - 73°	
Shock Position	+14°	
Maximum Weight Capacity	1600 / 850 lbs	(727/386 kg)
Minimum Operators Required		
Occupied Cot	2	
Unoccupied Cot	1	
Recommended Fastener System		
Floor Mount	Model 6370 or 6377	
Wall Mount	Model 6371	
Recommended Loading Height*	Up To 33"	(84 cm)

- **Height limit kit recommended for load heights less than 28" (71cm).**
High height wt recommended for load heights more than 32'(81 cm).
- 1 Height measured from bottom of mattress at seat section to ground level.**
- 2 Cot is weighed without mattress and restraints.**



Approved: 10/01/2011
Version: 3.1

Page 4 of 4

Approved by:


Charles Neal, D.O. Medical Director